

HR 8299

To allow for devices with a predetermined change control plan to be marketed without submitting a supplemental application or premarket notification if the changes to such devices are consistent with such plan.

Congress: 117 (2021–2023, Ended)

Chamber: House

Policy Area: Health

Introduced: Jul 7, 2022

Current Status: Referred to the House Committee on Energy and Commerce.

Latest Action: Referred to the House Committee on Energy and Commerce. (Jul 7, 2022)

Official Text: <https://www.congress.gov/bill/117th-congress/house-bill/8299>

Sponsor

Name: Rep. Bilirakis, Gus M. [R-FL-12]

Party: Republican • **State:** FL • **Chamber:** House

Cosponsors (7 total)

Cosponsor	Party / State	Role	Date Joined
Rep. O'Halleran, Tom [D-AZ-1]	D · AZ		Jul 7, 2022
Rep. Cárdenas, Tony [D-CA-29]	D · CA		Jul 28, 2022
Rep. Kuster, Ann M. [D-NH-2]	D · NH		Jul 28, 2022
Rep. Craig, Angie [D-MN-2]	D · MN		Jul 29, 2022
Rep. DeGette, Diana [D-CO-1]	D · CO		Sep 29, 2022
Rep. Rutherford, John H. [R-FL-4]	R · FL		Sep 29, 2022
Rep. Neguse, Joe [D-CO-2]	D · CO		Nov 14, 2022

Committee Activity

Committee	Chamber	Activity	Date
Energy and Commerce Committee	House	Referred To	Jul 7, 2022

Subjects & Policy Tags

Policy Area:

Health

Related Bills

Bill	Relationship	Last Action
117 S 4386	Identical bill	Jun 13, 2022: Read twice and referred to the Committee on Health, Education, Labor, and Pensions.

This bill allows changes to be made to medical devices without the need for supplemental applications or premarket notifications if the changes are made in accordance with a predetermined change control plan.

Actions Timeline

- **Jul 7, 2022:** Introduced in House
- **Jul 7, 2022:** Referred to the House Committee on Energy and Commerce.